


MEK-3DH
HEMATOLOGY CONTROL
2.0mL
ASSAY SHEET

LOT	B23YH
	2024-02-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420	MEK -6500 -6510	MEK -6450	MEK -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	20.5	20.5	20.5	20.5	20.08
RBC	10 ⁶ /μL	±0.25	5.79	5.79	5.79	5.79	5.81
HGB	g/dL	±0.7	19.2	19.1	19.2	19.1	19.14
HCT	%	±4.0	53.2	53.2	53.2	53.2	61.3
MCV	fL	±7.0	91.8	91.8	91.8	91.8	105.5
MCH	pg	±2.8	33.2	33.0	33.2	33.0	32.9
MCHC	g/dL	±3.0	36.1	35.9	36.1	35.9	31.2
PLT	10 ³ /μL	±70	592	592	592	592	597.7
MPV	fL	±3.0	6.2	8.5	6.2	8.5	9.1
RDW-CV *	%	±3.0	16.5	16.5	16.5	16.5	17.4
LY%	%	±10.0	16.4	16.4	16.4	16.4	14.41
MO%	%	±5.0	3.0	3.0	3.0	3.0	2.10
GR%	%	±10.0	80.6	80.6	80.6	80.6	83.49
LY	10 ³ /μL	±1.8	3.4	3.4	3.4	3.4	2.89
MO	10 ³ /μL	±0.7	0.6	0.6	0.6	0.6	0.42
GR	10 ³ /μL	±2.0	16.5	16.5	16.5	16.5	16.77

Manual Registration	
Lot	23Y
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>


NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU

TOKYO 161-8560 JAPAN



MEK-3DH



B23YH



2024-02-05



0694-901686N