


MK-RE2
HEMATOLOGY CONTROL
 for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE24Y-2
	2025-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	4.31	±2.69
RET	10⁶/μL	0.1516	±0.0946

2D Barcode Entry


• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
Manual Registration

Lot	24Y					
Level	2					
Expiration	1	M	5	D	2025	Y
Control	MKRE					

CE
 0123

IVD

NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN