


**MK-RE2** HEMATOLOGY CONTROL  
for reticulocyte 3.0mL

**ASSAY SHEET**

LOT	MKRE251-2
	2025-03-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	4.47	±2.53
RET	10 <sup>6</sup> /μL	0.1574	±0.0891

**2D Barcode Entry**



• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

Manual Registration						
Lot	251					
Level	2					
Expiration	3	M	5	D	2025	Y
Control	MKRE					



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN