


**MK-RE3**
**HEMATOLOGY CONTROL**  
 for reticulocyte

3.0mL

**ASSAY SHEET**

<b>LOT</b>	MKRE251-3
	2025-03-05
<b>OPEN</b>	14 DAYS

		<b>MEK-9200</b>	<b>RANGE</b>
<b>RET%</b>	<b>%</b>	<b>8.60</b>	<b>±3.40</b>
<b>RET</b>	<b>10<sup>6</sup>/μL</b>	<b>0.3019</b>	<b>±0.1194</b>

**2D Barcode Entry**


• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
**Manual Registration**

Lot	251					
Level	3					
Expiration	3	M	5	D	2025	Y
Control	MKRE					

**CE**  
 0123

**IVD**

**NIHON KOHDEN CORPORATION**  
 1-31-4 NISHIOCHIAI, SHINJUKU-KU  
 TOKYO 161-8560 JAPAN

0694-901710Y