


MEK-3DH	HEMATOLOGY CONTROL	2.0mL
----------------	---------------------------	--------------

ASSAY SHEET

LOT	B252H
	2025-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	21.6	21.6	20.86
RBC	10 ⁶ /μL	±0.25	5.92	5.92	5.91
HGB	g/dL	±0.7	19.3	19.3	19.15
HCT	%	±4.0	54.5	54.5	62.4
MCV	fL	±7.0	92.1	92.1	105.5
MCH	pg	±2.8	32.6	32.6	32.4
MCHC	g/dL	±3.0	35.4	35.4	30.7
PLT	10 ³ /μL	±70	554	554	554.3
MPV	fL	±3.0	5.2	7.9	8.3
RDW-CV *	%	±3.0	16.3	16.3	17.5
LY%	%	±10.0	15.0	15.0	11.30
MO%	%	±5.0	1.5	1.5	1.59
GR%	%	±10.0	83.5	83.5	87.11
LY	10 ³ /μL	±1.8	3.2	3.2	2.36
MO	10 ³ /μL	±0.7	0.3	0.3	0.33
GR	10 ³ /μL	±2.0	18.1	18.1	18.17

Manual Registration	
Lot	252
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>


NIHON KOHDEN CORPORATION

 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN


MEK-3DH



B252H



2025-05-05



0123



0694-901686S