


<b>MEK-3DN</b>	<b>HEMATOLOGY CONTROL</b>	<b>2.0mL</b>
----------------	---------------------------	--------------

**ASSAY SHEET**

<b>LOT</b>	<b>B252N</b>
	<b>2025-05-05</b>
<b>OPEN</b>	<b>14DAYS</b>

<b>MEK-1301,1302,1305 Barcode Entry</b>
---

The barcodes may be read in any order.  
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
<b>WBC</b>	10 <sup>3</sup> /μL	±0.8	<b>8.0</b>	<b>8.0</b>	<b>7.88</b>
<b>RBC</b>	10 <sup>6</sup> /μL	±0.20	<b>4.90</b>	<b>4.90</b>	<b>4.93</b>
<b>HGB</b>	g/dL	±0.5	<b>14.3</b>	<b>14.5</b>	<b>14.34</b>
<b>HCT</b>	%	±3.0	<b>41.8</b>	<b>41.8</b>	<b>48.0</b>
<b>MCV</b>	fL	±7.0	<b>85.3</b>	<b>85.3</b>	<b>97.4</b>
<b>MCH</b>	pg	±2.4	<b>29.2</b>	<b>29.6</b>	<b>29.1</b>
<b>MCHC</b>	g/dL	±3.0	<b>34.2</b>	<b>34.7</b>	<b>29.9</b>
<b>PLT</b>	10 <sup>3</sup> /μL	±50	<b>252</b>	<b>252</b>	<b>269.1</b>
<b>MPV</b>	fL	±2.0	<b>5.3</b>	<b>8.0</b>	<b>8.2</b>
<b>RDW-CV *</b>	%	±3.0	<b>16.7</b>	<b>16.7</b>	<b>16.7</b>
<b>LY%</b>	%	±10.0	<b>31.8</b>	<b>31.8</b>	<b>28.74</b>
<b>MO%</b>	%	±5.0	<b>3.9</b>	<b>3.9</b>	<b>4.14</b>
<b>GR%</b>	%	±10.0	<b>64.3</b>	<b>64.3</b>	<b>67.12</b>
<b>LY</b>	10 <sup>3</sup> /μL	±1.0	<b>2.5</b>	<b>2.5</b>	<b>2.26</b>
<b>MO</b>	10 <sup>3</sup> /μL	±0.4	<b>0.3</b>	<b>0.3</b>	<b>0.33</b>
<b>GR</b>	10 <sup>3</sup> /μL	±1.0	<b>5.2</b>	<b>5.2</b>	<b>5.29</b>

<b>MEK-1305</b>
<b>QC ESR</b>
<b>7±5 mm</b>

Manual Registration	
Lot	<b>252</b>
Level	<b>Normal</b>
Control	<b>MEK-3D</b>

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



**NIHON KOHDEN CORPORATION**  
 1-31-4 NISHIOCHIAI, SHINJUKU-KU  
 TOKYO 161-8560 JAPAN

