


MEK-5DL HEMATOLOGY CONTROL
for WBC 5 part differential 3.0mL

ASSAY SHEET

LOT	MEK5D252L
	2025-05-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry



		RANGE	MEK -7222	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.6	3.0	3.0	3.1	3.01
NE%	%	±15.0	27.5	27.5	27.5	29.93
LY%	%	±15.0	50.1	50.1	50.1	53.21
MO%	%	±5.0	17.5	17.5	17.5	12.68
EO%	%	±5.0	2.1	2.1	2.1	2.88
BA%	%	±4.0	2.8	2.8	2.8	1.30
NE	10 ³ /μL	±0.5	0.8	0.8	0.8	0.90
LY	10 ³ /μL	±0.5	1.5	1.5	1.6	1.60
MO	10 ³ /μL	±0.2	0.5	0.5	0.5	0.38
EO	10 ³ /μL	±0.2	0.1	0.1	0.1	0.09
BA	10 ³ /μL	±0.2	0.1	0.1	0.1	0.04
RBC	10 ⁶ /μL	±0.15	2.58	2.56	2.58	2.44
HGB	g/dL	±0.4	6.4	6.5	6.3	6.45
HCT	%	±2.0	21.7	20.8	20.5	22.2
MCV	fL	±7.0	84.0	81.4	79.3	91.1
MCH	pg	±2.4	24.8	25.4	24.4	26.4
MCHC	g/dL	±3.0	29.5	31.3	30.7	29.1
RDW-CV *	%	±4.0	14.9	16.3	15.2	15.2
PLT	10 ³ /μL	±30	53	60	54	55.3
MPV	fL	±3.0	3.9	5.1	4.2	6.9

Manual Registration	
Lot	252
Level	Low
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-7222,8222.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-5DL



252L



2025-05-05



0694-901690R