


MEK-5DN HEMATOLOGY CONTROL
for WBC 5 part differential 3.0mL

ASSAY SHEET

LOT	MEK5D252N
	2025-05-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry



		RANGE	MEK -7222	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.8	8.3	8.2	8.1	8.28
NE%	%	±15.0	43.8	43.8	43.8	46.44
LY%	%	±15.0	31.6	31.6	31.6	34.45
MO%	%	±5.0	18.4	18.4	18.4	12.86
EO%	%	±5.0	4.1	4.1	4.1	4.80
BA%	%	±4.0	2.1	2.1	2.1	1.45
NE	10 ³ /μL	±1.5	3.7	3.6	3.5	3.85
LY	10 ³ /μL	±1.5	2.6	2.6	2.6	2.85
MO	10 ³ /μL	±0.5	1.5	1.5	1.5	1.06
EO	10 ³ /μL	±0.5	0.3	0.3	0.3	0.40
BA	10 ³ /μL	±0.4	0.2	0.2	0.2	0.12
RBC	10 ⁶ /μL	±0.20	4.50	4.43	4.48	4.46
HGB	g/dL	±0.5	13.4	13.3	13.2	13.53
HCT	%	±3.0	42.5	40.7	40.0	46.7
MCV	fL	±6.0	94.4	91.8	89.3	104.6
MCH	pg	±2.8	29.8	30.0	29.5	30.3
MCHC	g/dL	±3.0	31.5	32.7	33.0	29.0
RDW-CV *	%	±4.0	13.5	14.2	13.7	13.9
PLT	10 ³ /μL	±50	283	301	276	267.8
MPV	fL	±2.0	3.0	4.9	3.2	6.5

Manual Registration	
Lot	252
Level	Normal
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-7222,8222.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901691R