


**MK-RE3**

**HEMATOLOGY CONTROL**  
for reticulocyte

3.0mL

**ASSAY SHEET**

LOT	MKRE253-3
	2025-05-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	8.70	±3.30
RET	10 <sup>6</sup> /μL	0.3087	±0.1171

**2D Barcode Entry**



- If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

**Manual Registration**

Lot	253					
Level	3					
Expiration	5	M	5	D	2025	Y
Control	MKRE					



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN