

MEK-5DL
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D255L
	2025-08-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7222	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.6	3.1	3.1	3.1	3.08
NE%	%	±15.0	45.8	45.8	45.8	47.14
LY%	%	±15.0	37.3	37.3	37.3	40.28
MO%	%	±5.0	13.3	13.3	13.3	9.64
EO%	%	±5.0	1.0	1.0	1.0	1.20
BA%	%	±4.0	2.6	2.6	2.6	1.74
IG% **	%	±4.0	-	-	-	0.40
NE	10 ³ /μL	±0.5	1.4	1.4	1.4	1.45
LY	10 ³ /μL	±0.5	1.2	1.2	1.2	1.24
MO	10 ³ /μL	±0.2	0.4	0.4	0.4	0.30
EO	10 ³ /μL	±0.2	0.0	0.0	0.0	0.04
BA	10 ³ /μL	±0.2	0.1	0.1	0.1	0.05
IG **	10 ³ /μL	±0.2	-	-	-	0.01
RBC	10 ⁶ /μL	±0.15	2.54	2.52	2.55	2.41
HGB	g/dL	±0.4	6.3	6.3	6.2	6.33
HCT	%	±2.0	21.4	20.6	20.3	22.1
MCV	fL	±7.0	84.3	81.7	79.6	91.5
MCH	pg	±2.4	24.8	25.0	24.3	26.3
MCHC	g/dL	±3.0	29.4	30.6	30.5	28.6
RDW-CV *	%	±4.0	15.2	16.6	15.5	15.8
PLT	10 ³ /μL	±30	53	60	54	54.6
MPV	fL	±3.0	4.3	5.3	4.5	7.1

Manual Registration

Lot	255
Level	Low
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-7222,8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DL



255L



2025-08-05



0694-901690S(01)