

MEK-3DN

HEMATOLOGY CONTROL

2.0mL

ASSAY SHEET

LOT	B258N
	2025-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	7.9	7.9	7.95
RBC	10 ⁶ /μL	±0.20	4.78	4.78	4.73
HGB	g/dL	±0.5	13.8	14.0	14.09
HCT	%	±3.0	41.3	41.3	46.7
MCV	fL	±7.0	86.3	86.3	98.8
MCH	pg	±2.4	28.9	29.3	29.8
MCHC	g/dL	±3.0	33.4	33.9	30.2
PLT	10 ³ /μL	±50	242	242	263.1
MPV	fL	±2.0	5.5	8.1	8.2
RDW-CV *	%	±3.0	16.8	16.8	16.3
LY%	%	±10.0	31.9	31.9	31.03
MO%	%	±5.0	2.7	2.7	2.93
GR%	%	±10.0	65.4	65.4	66.04
LY	10 ³ /μL	±1.0	2.5	2.5	2.47
MO	10 ³ /μL	±0.4	0.2	0.2	0.23
GR	10 ³ /μL	±1.0	5.2	5.2	5.25

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.

MEK-1305
QC ESR
8±5 mm

Manual Registration

Lot	258
Level	Normal
Control	MEK-3D


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-3DN



B258N



2025-11-05



0123



0694-901685U