

MEK-5DH
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D258H
	2025-11-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±2.2	24.6	24.2	25.63
NE%	%	±10.0	74.2	74.2	74.72
LY%	%	±10.0	13.4	13.4	15.03
MO%	%	±5.0	7.8	7.8	5.67
EO%	%	±5.0	3.7	3.7	3.73
BA%	%	±4.0	0.9	0.9	0.85
IG% **	%	±4.0	-	-	0.45
NE	10 ³ /μL	±2.8	18.3	18.0	19.15
LY	10 ³ /μL	±2.8	3.3	3.2	3.85
MO	10 ³ /μL	±1.0	1.9	1.9	1.45
EO	10 ³ /μL	±1.0	0.9	0.9	0.96
BA	10 ³ /μL	±0.8	0.2	0.2	0.22
IG **	10 ³ /μL	±0.8	-	-	0.12
RBC	10 ⁶ /μL	±0.25	5.26	5.25	5.32
HGB	g/dL	±0.7	16.3	16.4	16.61
HCT	%	±4.0	51.2	49.8	59.0
MCV	fL	±8.0	97.4	94.9	110.9
MCH	pg	±2.8	31.0	31.2	31.2
MCHC	g/dL	±3.0	31.8	32.9	28.2
RDW-CV *	%	±4.0	14.2	13.6	14.6
PLT	10 ³ /μL	±70	534	541	538.1
MPV	fL	±4.0	8.2	8.1	9.7

Manual Registration	
Lot	258
Level	High
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DH



258H



2025-11-05



0694-901692T