

MEK-5DL
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D258L
	2025-11-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.6	3.1	3.1	3.11
NE%	%	±15.0	34.4	34.4	34.76
LY%	%	±15.0	45.3	45.3	49.62
MO%	%	±5.0	14.9	14.9	10.71
EO%	%	±5.0	2.8	2.8	3.34
BA%	%	±4.0	2.6	2.6	1.57
IG% **	%	±4.0	-	-	1.00
NE	10 ³ /μL	±0.5	1.0	1.0	1.09
LY	10 ³ /μL	±0.5	1.4	1.4	1.54
MO	10 ³ /μL	±0.2	0.5	0.5	0.33
EO	10 ³ /μL	±0.2	0.1	0.1	0.10
BA	10 ³ /μL	±0.2	0.1	0.1	0.05
IG **	10 ³ /μL	±0.2	-	-	0.03
RBC	10 ⁶ /μL	±0.15	2.74	2.77	2.64
HGB	g/dL	±0.4	6.7	6.6	6.81
HCT	%	±2.0	21.9	21.6	23.6
MCV	fL	±7.0	79.9	77.8	89.3
MCH	pg	±2.4	24.5	23.8	25.8
MCHC	g/dL	±3.0	30.6	30.6	28.9
RDW-CV *	%	±4.0	16.2	15.2	15.1
PLT	10 ³ /μL	±30	60	53	55.4
MPV	fL	±3.0	5.4	4.7	6.9

Manual Registration	
Lot	258
Level	Low
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DL



258L



2025-11-05



0694-901690T