

MEK-5DN
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D258N
	2025-11-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.8	8.0	7.8	8.08
NE%	%	±15.0	54.4	54.4	54.58
LY%	%	±15.0	28.5	28.5	31.66
MO%	%	±5.0	11.9	11.9	8.58
EO%	%	±5.0	3.4	3.4	3.65
BA%	%	±4.0	1.8	1.8	1.53
IG% **	%	±4.0	-	-	0.55
NE	10 ³ /μL	±1.5	4.3	4.3	4.42
LY	10 ³ /μL	±1.5	2.3	2.2	2.56
MO	10 ³ /μL	±0.5	1.0	0.9	0.69
EO	10 ³ /μL	±0.5	0.3	0.3	0.29
BA	10 ³ /μL	±0.4	0.1	0.1	0.12
IG **	10 ³ /μL	±0.4	-	-	0.04
RBC	10 ⁶ /μL	±0.20	4.54	4.59	4.54
HGB	g/dL	±0.5	13.3	13.2	13.69
HCT	%	±3.0	41.2	40.5	46.9
MCV	fL	±6.0	90.8	88.3	103.4
MCH	pg	±2.8	29.3	28.8	30.2
MCHC	g/dL	±3.0	32.3	32.6	29.2
RDW-CV *	%	±4.0	14.5	13.9	14.2
PLT	10 ³ /μL	±50	287	264	259.2
MPV	fL	±2.0	5.1	3.5	6.6

Manual Registration	
Lot	258
Level	Normal
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DN



258N



2025-11-05



0694-901691T