

MEK-CAL

HEMATOLOGY CALIBRATOR

2.0mL

ASSAY SHEET

LOT	PLUS258
	2025-09-05
OPEN	7DAYS

MEK-9100,9200
2D Barcode Entry

MEK-1301,1302,1303,1305
Barcode Entry


MEK-CAL



PLUS258



2025-09-05



0123



0694-901832C

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -7300	MEK -8222	MEK -9100 -9200	MEK -1301 -1302 -1303 -1305
WBC	10 ³ /μL	9.1	9.1	9.2	9.1	9.53	9.25
RBC	10 ⁶ /μL	4.77	4.77	4.77	4.73	4.74	4.72
HGB	g/dL	13.7	13.8	13.6	13.6	13.99	13.89
HCT	%	40.7	40.7	41.4	40.3	46.7	46.7
MCV	fL	85.4	85.4	86.8	85.1	98.5	99.0
MCH	pg	28.7	28.9	28.5	28.8	29.5	29.4
MCHC	g/dL	33.7	33.9	32.9	33.7	30.0	29.7
PLT	10 ³ /μL	226	226	227	226	232.1	240.1
MPV	fL	6.0	8.4	7.2	6.0	8.5	8.6
RDW-CV *	%	15.8	15.8	16.0	15.0	15.9	15.8

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450,8222.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN