



Wireless Module

QI-101V

General

The QI-101V wireless module is a wireless module for the TEC-1000 series defibrillators.

The wireless module is installed in the defibrillator to transmit and receive data via Bluetooth and wireless LAN.

Safety Information

- ⚠ WARNING** A warning alerts the user to possible injury or death associated with the use or misuse of the instrument.
- ⚠ CAUTION** A caution alerts the user to possible injury or problems with the instrument associated with its use or misuse such as instrument malfunction, instrument failure, damage to the instrument, or damage to other property.

Pay attention to all safety information in the Operator's Manual or Installation Guide.

⚠ WARNING

Connect the wireless module only to the specified instrument and follow the specified procedure. Failure to follow this warning may result in electrical shock or injury to the patient and operator.

⚠ CAUTION

Before connecting or disconnecting instruments, make sure that each instrument is turned off, the power cord is disconnected from the AC socket, and the battery pack is not installed. Otherwise, the patient or operator may receive electrical shock or injury.

This Safety and Performance Information is an extract from the general and safety information sections of the most recent edition of Operator's Manual or Installation Guide. Therefore, the contents of your Operator's Manual or Installation Guide may differ from those of this Safety and Performance Information. For detailed operating procedures, follow the instructions of your Operator's Manual or Installation Guide.



Manufacturer

NIHON KOHDEN CORPORATION
1-31-4 Nishiochiai, Shinjuku-ku,
Tokyo 161-8560, Japan
Phone +81 3-5996-8041
<https://www.nihonkohden.com/>



European Representative

NIHON KOHDEN EUROPE GmbH
Raiffeisenstrasse 10, 61191 Rosbach, Germany
Phone +49 6003-827-0 Fax +49 6003-827-599