

MEK-5DH
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D25YH
	2026-02-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±2.2	24.4	23.9	24.98
NE%	%	±10.0	75.2	75.2	75.08
LY%	%	±10.0	12.3	12.3	14.50
MO%	%	±5.0	7.5	7.5	5.81
EO%	%	±5.0	4.1	4.1	3.46
BA%	%	±4.0	0.9	0.9	1.15
IG% **	%	±4.0	-	-	0.34
NE	10 ³ /μL	±2.8	18.4	18.0	18.76
LY	10 ³ /μL	±2.8	3.0	2.9	3.62
MO	10 ³ /μL	±1.0	1.8	1.8	1.45
EO	10 ³ /μL	±1.0	1.0	1.0	0.86
BA	10 ³ /μL	±0.8	0.2	0.2	0.29
IG **	10 ³ /μL	±0.8	-	-	0.08
RBC	10 ⁶ /μL	±0.25	5.48	5.47	5.54
HGB	g/dL	±0.7	17.4	17.6	17.68
HCT	%	±4.0	54.4	52.9	62.3
MCV	fL	±8.0	99.3	96.7	112.4
MCH	pg	±2.8	31.8	32.2	31.9
MCHC	g/dL	±3.0	32.0	33.3	28.4
RDW-CV *	%	±4.0	14.5	13.8	15.1
PLT	10 ³ /μL	±70	548	555	560.3
MPV	fL	±4.0	7.1	6.0	8.6

Manual Registration	
Lot	25Y
Level	High
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DH



25YH



2026-02-05



0123



0694-901692U