

MEK-5DL
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D25YL
	2026-02-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.6	2.9	3.0	2.93
NE%	%	±15.0	49.7	49.7	47.17
LY%	%	±15.0	36.8	36.8	43.12
MO%	%	±5.0	9.7	9.7	6.79
EO%	%	±5.0	1.8	1.8	1.41
BA%	%	±4.0	2.0	2.0	1.51
IG% **	%	±4.0	-	-	1.74
NE	10 ³ /μL	±0.5	1.3	1.4	1.39
LY	10 ³ /μL	±0.5	1.1	1.1	1.26
MO	10 ³ /μL	±0.2	0.3	0.3	0.20
EO	10 ³ /μL	±0.2	0.1	0.1	0.04
BA	10 ³ /μL	±0.2	0.1	0.1	0.04
IG **	10 ³ /μL	±0.2	-	-	0.05
RBC	10 ⁶ /μL	±0.15	2.58	2.61	2.45
HGB	g/dL	±0.4	6.5	6.4	6.55
HCT	%	±2.0	21.2	20.9	22.5
MCV	fL	±7.0	82.2	80.1	91.8
MCH	pg	±2.4	25.2	24.5	26.7
MCHC	g/dL	±3.0	30.7	30.6	29.1
RDW-CV *	%	±4.0	16.3	15.3	15.3
PLT	10 ³ /μL	±30	56	50	49.3
MPV	fL	±3.0	5.3	4.5	6.8

Manual Registration	
Lot	25Y
Level	Low
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DL



25YL



2026-02-05



0123



0694-901690U