

MEK-5DN
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D25YN
	2026-02-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±0.8	8.1	8.0	8.26
NE%	%	±15.0	57.3	57.3	56.82
LY%	%	±15.0	26.9	26.9	29.76
MO%	%	±5.0	10.8	10.8	8.33
EO%	%	±5.0	3.3	3.3	3.77
BA%	%	±4.0	1.7	1.7	1.32
IG% **	%	±4.0	-	-	0.38
NE	10 ³ /μL	±1.5	4.6	4.5	4.69
LY	10 ³ /μL	±1.5	2.2	2.2	2.46
MO	10 ³ /μL	±0.5	0.9	0.9	0.69
EO	10 ³ /μL	±0.5	0.3	0.3	0.31
BA	10 ³ /μL	±0.4	0.1	0.1	0.11
IG **	10 ³ /μL	±0.4	-	-	0.03
RBC	10 ⁶ /μL	±0.20	4.68	4.73	4.70
HGB	g/dL	±0.5	13.5	13.4	13.74
HCT	%	±3.0	41.8	41.1	47.8
MCV	fL	±6.0	89.3	86.9	101.8
MCH	pg	±2.8	28.8	28.3	29.2
MCHC	g/dL	±3.0	32.3	32.6	28.7
RDW-CV *	%	±4.0	15.3	14.5	15.1
PLT	10 ³ /μL	±50	292	269	266.9
MPV	fL	±2.0	5.4	4.0	7.0

Manual Registration	
Lot	25Y
Level	Normal
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.



<https://www.nihonkohden.com/spi/index.html>


NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DN



25YN



2026-02-05



0123



0694-901691U