


MEK-5DH
HEMATOLOGY CONTROL
for WBC 5 part differential
3.0mL
ASSAY SHEET

LOT	MEK5D265H
	2026-08-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10³/μL	±2.2	24.2	23.7	25.04
NE%	%	±10.0	82.8	82.8	83.90
LY%	%	±10.0	7.8	7.8	8.71
MO%	%	±5.0	7.1	7.1	4.57
EO%	%	±5.0	1.8	1.8	2.15
BA%	%	±4.0	0.5	0.5	0.67
IG% **	%	±4.0	-	-	0.79
NE	10³/μL	±2.8	20.1	19.7	21.01
LY	10³/μL	±2.8	1.9	1.8	2.18
MO	10³/μL	±1.0	1.7	1.7	1.14
EO	10³/μL	±1.0	0.4	0.4	0.54
BA	10³/μL	±0.8	0.1	0.1	0.17
IG **	10³/μL	±0.8	-	-	0.20
RBC	10⁶/μL	±0.25	5.24	5.23	5.31
HGB	g/dL	±0.7	16.7	16.8	16.94
HCT	%	±4.0	51.8	50.3	59.0
MCV	fL	±8.0	98.8	96.2	111.1
MCH	pg	±2.8	31.9	32.1	31.9
MCHC	g/dL	±3.0	32.2	33.4	28.7
RDW-CV *	%	±4.0	14.6	13.9	14.8
PLT	10³/μL	±70	555	562	548.3
MPV	fL	±4.0	6.5	4.8	7.7

Manual Registration	
Lot	265
Level	High
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
 Only software versions 02-01 and later can be used.



<https://www.nihonkohden.com/spi/index.html>


NIHON KOHDEN CORPORATION

 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN


MEK-5DH



265H



2026-08-05



0694-901692W