


MEK-5DL	HEMATOLOGY CONTROL for WBC 5 part differential	3.0mL
----------------	--	-------

ASSAY SHEET

LOT	MEK5D265L
	2026-08-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry



		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10³/μL	±0.6	3.0	3.0	2.96
NE%	%	±15.0	46.0	46.0	44.36
LY%	%	±15.0	37.9	37.9	43.48
MO%	%	±5.0	8.7	8.7	6.20
EO%	%	±5.0	4.8	4.8	3.87
BA%	%	±4.0	2.6	2.6	2.09
IG% **	%	±4.0	-	-	0.46
NE	10³/μL	±0.5	1.4	1.4	1.32
LY	10³/μL	±0.5	1.1	1.1	1.29
MO	10³/μL	±0.2	0.3	0.3	0.18
EO	10³/μL	±0.2	0.1	0.1	0.11
BA	10³/μL	±0.2	0.1	0.1	0.06
IG **	10³/μL	±0.2	-	-	0.01
RBC	10⁶/μL	±0.15	2.62	2.65	2.54
HGB	g/dL	±0.4	6.8	6.6	6.79
HCT	%	±2.0	21.8	21.5	23.4
MCV	fL	±7.0	83.1	81.0	92.2
MCH	pg	±2.4	26.0	24.9	26.7
MCHC	g/dL	±3.0	31.2	30.7	29.0
RDW-CV *	%	±4.0	15.8	14.9	14.5
PLT	10³/μL	±30	54	49	50.1
MPV	fL	±3.0	5.3	4.6	6.9

Manual Registration	
Lot	265
Level	Low
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-5DL



265L



2026-08-05



0694-901690W