


MK-3CN HEMATOLOGY CONTROL+CRP 2.5mL

ASSAY SHEET

LOT	MK3C265N
	2026-08-05
OPEN	14DAYS

MEK-1303 Barcode Entry

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		MEK-1303	RANGE
WBC	10 ³ /μL	7.57	± 0.80
RBC	10 ⁶ /μL	4.76	± 0.20
HGB	g/dL	14.28	± 0.50
HCT	%	47.1	± 3.0
MCV	fL	98.9	± 9.0
MCH	pg	30.0	± 2.4
MCHC	g/dL	30.3	± 3.0
PLT	10 ³ /μL	263.9	± 50.0
MPV	fL	8.0	± 2.0
RDW-CV	%	16.8	± 3.0
LY%	%	30.15	± 10.00
MO%	%	3.19	± 5.00
GR%	%	66.66	± 10.00
LY	10 ³ /μL	2.28	± 1.00
MO	10 ³ /μL	0.24	± 0.40
GR	10 ³ /μL	5.05	± 1.00
CRP	Precision	≤ 10.0% (18.70-34.70 mg/L)	

Manual Registration

Lot	265
Level	Normal
Control	MK-3C


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN


0123



0694-901766N