


**MK-RE2**
**HEMATOLOGY CONTROL**  
 for reticulocyte

3.0mL

**ASSAY SHEET**

<b>LOT</b>	MKRE265-2
	2026-07-05
<b>OPEN</b>	14 DAYS

		MEK-9200	RANGE
<b>RET%</b>	<b>%</b>	<b>4.40</b>	<b>±2.60</b>
<b>RET</b>	<b>10<sup>6</sup>/μL</b>	<b>0.1549</b>	<b>±0.0915</b>

**2D Barcode Entry**


• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
**Manual Registration**

Lot	265					
Level	2					
Expiration	7	M	5	D	2026	Y
Control	MKRE					

  
 0123



**NIHON KOHDEN CORPORATION**  
 1-31-4 NISHIOCHIAI, SHINJUKU-KU  
 TOKYO 161-8560 JAPAN