




**MK-RE3**

**HEMATOLOGY CONTROL**  
for reticulocyte

3.0mL

**ASSAY SHEET**

 LOT	MKRE265-3
	2026-07-05
 OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	9.02	±2.98
RET	10 <sup>6</sup> /μL	0.3181	±0.1051

**2D Barcode Entry**



• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

**Manual Registration**

Lot	265					
Level	3					
Expiration	7	M	5	D	2026	Y
Control	MKRE					

  
0123





**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN