

MEK-5DH
HEMATOLOGY CONTROL
for WBC 5 part differential

3.0mL

ASSAY SHEET

LOT	MEK5D268H
	2026-11-05
OPEN	14DAYS

MEK-9100,9200
2D Barcode Entry


		RANGE	MEK -7300	MEK -8222	MEK -9100 -9200
WBC	10 ³ /μL	±2.2	25.2	24.7	26.32
NE%	%	±10.0	73.3	73.3	76.19
LY%	%	±10.0	14.4	14.4	15.11
MO%	%	±5.0	9.1	9.1	5.07
EO%	%	±5.0	2.4	2.4	2.82
BA%	%	±4.0	0.8	0.8	0.81
IG% **	%	±4.0	-	-	0.97
NE	10 ³ /μL	±2.8	18.5	18.1	20.06
LY	10 ³ /μL	±2.8	3.6	3.6	3.98
MO	10 ³ /μL	±1.0	2.3	2.2	1.33
EO	10 ³ /μL	±1.0	0.6	0.6	0.74
BA	10 ³ /μL	±0.8	0.2	0.2	0.21
IG **	10 ³ /μL	±0.8	-	-	0.26
RBC	10 ⁶ /μL	±0.25	5.31	5.30	5.42
HGB	g/dL	±0.7	16.3	16.4	16.48
HCT	%	±4.0	50.1	48.7	57.8
MCV	fL	±8.0	94.3	91.9	106.6
MCH	pg	±2.8	30.7	30.9	30.4
MCHC	g/dL	±3.0	32.5	33.7	28.5
RDW-CV *	%	±4.0	14.5	13.9	15.2
PLT	10 ³ /μL	±70	513	520	523.4
MPV	fL	±4.0	7.7	7.1	9.2

Manual Registration	
Lot	268
Level	High
Control	MEK5D

* RDW-CV is displayed as RDW on MEK-8222.

** MEK-9200 only
Only software versions 02-01 and later can be used.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-5DH



268H



2026-11-05



0123



0694-901692X