

MEK-CAL

HEMATOLOGY CALIBRATOR

2.0mL

ASSAY SHEET

LOT	PLUS268
	2026-09-05
OPEN	7DAYS

MEK-9100,9200
2D Barcode Entry

MEK-1301,1302,1303,1305
Barcode Entry


MEK-CAL



PLUS268



2026-09-05



0123



0694-9018320

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -7300	MEK -8222	MEK -9100 -9200	MEK -1301 -1302 -1303 -1305
WBC	10 ³ /μL	9.1	9.1	9.1	9.0	9.40	9.16
RBC	10 ⁶ /μL	4.76	4.76	4.71	4.71	4.74	4.76
HGB	g/dL	13.7	13.8	13.6	13.5	13.90	13.86
HCT	%	40.7	40.7	40.6	39.5	46.1	46.5
MCV	fL	85.6	85.6	86.1	83.9	97.3	97.6
MCH	pg	28.8	29.0	28.9	28.7	29.3	29.1
MCHC	g/dL	33.7	33.9	33.5	34.2	30.2	29.8
PLT	10 ³ /μL	211	211	215	214	224.2	227.7
MPV	fL	5.7	8.2	7.1	5.7	8.3	8.4
RDW-CV *	%	16.2	16.2	16.1	15.1	16.1	15.8

MEK -7300 -8222 -9100 -9200	
FS	166
FL	89
SD	103
FS THR.	100

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450,8222.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN