

MK-3CL HEMATOLOGY CONTROL+CRP 2.5mL

ASSAY SHEET

LOT	MK3C268L
	2026-11-05
OPEN	14DAYS

MEK-1303 Barcode Entry

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		MEK-1303	RANGE
WBC	10 ³ /μL	1.98	± 0.60
RBC	10 ⁶ /μL	2.32	± 0.15
HGB	g/dL	6.15	± 0.40
HCT	%	20.4	± 2.0
MCV	fL	88.1	± 9.0
MCH	pg	26.5	± 2.4
MCHC	g/dL	30.1	± 3.0
PLT	10 ³ /μL	71.2	± 30.0
MPV	fL	8.2	± 3.0
RDW-CV	%	16.7	± 3.0
LY%	%	57.14	± 10.00
MO%	%	6.10	± 5.00
GR%	%	36.76	± 10.00
LY	10 ³ /μL	1.13	± 0.50
MO	10 ³ /μL	0.12	± 0.20
GR	10 ³ /μL	0.73	± 0.50
CRP	Precision	≤ 10.0%	
		(2.30-12.70 mg/L)	

Manual Registration

Lot	268
Level	Low
Control	MK-3C


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

CE
0123

IVD

0694-9017650