

MK-3CN HEMATOLOGY CONTROL+CRP

2.5mL

ASSAY SHEET

LOT	MK3C268N
	2026-11-05
OPEN	14DAYS

MEK-1303 Barcode Entry

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		MEK-1303	RANGE
WBC	10 ³ /μL	7.41	± 0.80
RBC	10 ⁶ /μL	4.81	± 0.20
HGB	g/dL	13.90	± 0.50
HCT	%	46.0	± 3.0
MCV	fL	95.7	± 7.0
MCH	pg	28.9	± 2.4
MCHC	g/dL	30.2	± 3.0
PLT	10 ³ /μL	244.4	± 50.0
MPV	fL	8.0	± 2.0
RDW-CV	%	16.7	± 3.0
LY%	%	30.26	± 10.00
MO%	%	2.76	± 5.00
GR%	%	66.98	± 10.00
LY	10 ³ /μL	2.24	± 1.00
MO	10 ³ /μL	0.20	± 0.40
GR	10 ³ /μL	4.97	± 1.00
CRP	Precision	≤ 10.0%	
		(18.80-35.00 mg/L)	

Manual Registration

Lot	268
Level	Normal
Control	MK-3C


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

CE
0123

IVD

0694-9017660