

MK-RE1
HEMATOLOGY CONTROL
for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE269-1
	2026-11-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	1.70	±1.30
RET	10 ⁶ /μL	0.0597	±0.0457

2D Barcode Entry


- If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi>
Manual Registration

Lot	269					
Level	1					
Expiration	11	M	5	D	2026	Y
Control	MKRE					

CE
0123

IVD


NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

0694-901848H