ASSAY SHEET

NIHON KOHDEN

MEK-3DH HEMATOLOGY CONTROL

MEK-3DH

LOT	B26YH	
	2027-02-05	
OPEN	14DAYS	

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305	RANGE
WBC	$10^3/\mu L$	21.2	21.2	20.84	± 2.2
RBC	$10^6/\mu L$	6.00	6.00	6.06	± 0.25
HGB	g/dL	19.2	19.1	19.61	± 0.7
HCT	%	55.1	55.1	63.9	± 4.0
MCV	fL	91.9	91.9	105.5	± 7.0
MCH	pg	32.0	31.8	32.4	± 2.8
MCHC	g/dL	34.8	34.7	30.7	± 3.0
PLT	$10^3/\mu L$	521	521	528.5	± 70
MPV	fL	5.0	7.9	8.0	± 3.0
RDW-CV *1	%	16.5	16.5	17.4	± 3.0
LY%	%	16.3	16.3	14.47	± 10.0
MO%	%	1.7	1.7	1.66	± 5.0
GR%	%	82.0	82.0	83.87	± 10.0
LY	$10^3/\mu L$	3.5	3.5	3.02	± 1.8
MO	$10^3/\mu L$	0.4	0.4	0.35	± 0.7
GR	$10^3/\mu L$	17.3	17.3	17.47	± 2.0

*1 RDW-CV is displayed as RDW on MEK-6400, 6410, 6420, 6450.

Manual Registration	
Lot	26Y
Level	High
Expiration Date	2027 Y 2 M 5 D
Control	MEK-3D

MEK-1301, 1302, 1305 Barcode Entry

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1

2

3

MEK-1301, 1302, 1305 2D Barcode Entry

https://www.nihonkohden.com/spi

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

CE 0123 IVD

0694-901686Z