ASSAY SHEET

NIHON KOHDEN

MEK-3DL HEMATOLOGY CONTROL

MEK-3DL

LOT	B26YL	
	2027-02-05	
OPEN	14DAYS	

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305	RANGE
WBC	$10^3/\mu L$	2.1	2.1	2.09	± 0.6
RBC	$10^6/\mu L$	2.39	2.39	2.32	± 0.15
HGB	g/dL	6.0	6.1	6.14	± 0.4
HCT	%	18.5	18.5	20.4	± 2.0
MCV	fL	77.5	77.5	88.0	± 9.0
MCH	pg	25.1	25.5	26.5	± 2.4
MCHC	g/dL	32.4	33.0	30.1	± 3.0
PLT	$10^3/\mu L$	61	61	72.1	± 30
MPV	fL	5.9	8.4	8.3	± 3.0
RDW-CV *1	%	18.1	18.1	15.8	± 3.0
LY%	%	56.7	56.7	57.02	± 10.0
MO%	%	9.5	9.5	5.56	± 5.0
GR%	%	33.8	33.8	37.42	± 10.0
LY	$10^3/\mu L$	1.2	1.2	1.19	± 0.5
MO	$10^3/\mu L$	0.2	0.2	0.12	± 0.2
GR	$10^3/\mu L$	0.7	0.7	0.78	± 0.5
QC ESR *2	mm	-	-	48	± 17

*1 RDW-CV is displayed as RDW on MEK-6400, 6410, 6420, 6450.

*2 MEK-1305 only.

Manual Registration	
Lot	26Y
Level	Low
Expiration Date	2027 Y 2 M 5 D
Control	MEK-3D

MEK-1301, 1302, 1305 Barcode Entry

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1

2

3

MEK-1301, 1302, 1305 2D Barcode Entry

https://www.nihonkohden.com/spi

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

CE 0123 IVD

0694-901684Z