ASSAY SHEET

NIHON KOHDEN

MEK-3DN HEMATOLOGY CONTROL

MEK-3DN

LOT	B26YN
⌛	2027-02-05
OPEN	14DAYS

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305	RANGE
WBC	10^3/μL	7.9	7.9	7.81	± 0.8
RBC	10^6/μL	4.79	4.79	4.76	± 0.20
HGB	g/dL	13.9	14.0	14.15	± 0.5
HCT	%	41.3	41.3	46.8	± 3.0
MCV	fL	86.2	86.2	98.4	± 7.0
MCH	pg	29.0	29.2	29.7	± 2.4
MCHC	g/dL	33.7	33.9	30.2	± 3.0
PLT	10^3/μL	245	245	260.8	± 50
MPV	fL	5.4	8.1	8.2	± 2.0
RDW-CV *1	%	16.7	16.7	16.3	± 3.0
LY%	%	30.2	30.2	28.81	± 10.0
MO%	%	4.6	4.6	4.15	± 5.0
GR%	%	65.2	65.2	67.04	± 10.0
LY	10^3/μL	2.4	2.4	2.25	± 1.0
MO	10^3/μL	0.4	0.4	0.32	± 0.4
GR	10^3/μL	5.1	5.1	5.24	± 1.0
QC ESR *2	mm	-	-	9	± 5

*1 RDW-CV is displayed as RDW on MEK-6400, 6410, 6420, 6450.

*2 MEK-1305 only.

Manual Registration	
Lot	26Y
Level	Normal
Expiration Date	2027 Y 2 M 5 D
Control	MEK-3D

MEK-1301, 1302, 1305 Barcode Entry

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1

2

3

MEK-1301, 1302, 1305 2D Barcode Entry

https://www.nihonkohden.com/spi

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

CE 0123 IVD

0694-901685Z