NIHON KOHDEN

MK-RE2 HEMATOLOGY CONTROL for reticulocyte 3.0mL

ASSAY SHEET

LOT	MKRE26Y-2
	2027-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	4.45	±2.55
RET	$10^6/\mu L$	0.1591	±0.0912

2D Barcode Entry

- **If RBC value is under 3.00 [$\times 10^6/\mu L$], please refrain from using. It may have deteriorated or hemolyzed, which affects the measurement accuracy.**

https://www.nihonkohden.com/spi

Manual Registration						
Lot	26Y					
Level	2					
Expiration	1	M	5	D	2027	Y
Control	MKRE					

CE 0123 IVD

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

0694-901849I